



dentair

GZ Institute of Microbiology

**TEST REPORTS**

All tests conducted with representative sample units

GZ INSTITUTE OF MICROBIOLOGY

# Sterilization Rate

TEST REPORT

GZ INSTITUTE OF MICROBIOLOGY  
TEST REPORT

Date Received: Jul. 02, 2020

Date Analyzed: Jul. 02, 2020

Name of Sample	Air Purifier	Source of Sample	Delivery
Applicant	Bryant Medical LTD	Client	Chen Meifang
Manufacturer	---	Brand	dentair®
Type and Specification	FOZKYGB-03	Quantity of Sample	1PC
Date of Production	---	State of Sample	Machine
Batch Number	---	Packing of Sample	In box
Sample Picture			
Standard and Methods	<ol style="list-style-type: none"> <li>1. &lt;Technical Standard For Disinfection&gt;2002-2.1.3 Air disinfection effect evaluation test</li> <li>2. Referring to T/GIEHA 009-2018 The method for removing allergens of air cleaner</li> </ol>		
Items of Analysis	<ol style="list-style-type: none"> <li>1. Killing Rate (<i>Staphylococcus aureus</i> ATCC 6538, <i>Escherichia coli</i> 8099, <i>Klebsiella pneumoniae</i> ATCC 4352)</li> <li>2. *Mite Antigen Removal Rate (<i>Dust mite Der f 1</i>)</li> </ol>		
Remarks	---		

\*\*\*To be continued\*\*\*

GZ INSTITUTE OF MICROBIOLOGY  
TEST REPORT

Date Received: Jul. 02, 2020  
Date Analyzed: Jul. 02, 2020

**Method for Testing Air Disinfection:**

1. Test Equipments
  - 1) Test microorganism: *Staphylococcus aureus*, *Escherichia coli*, *Klebsiella pneumoniae*
  - 2) Microbial aerosol generator: TK-3
  - 3) Culture media: NA
  - 4) Sampling equipment: six-stage sieve sampler
2. Test Conditions
  - 1) The volume of the test chamber: 30 m<sup>3</sup>
  - 2) Environment temperature: (20~25) °C
  - 3) Environment humidity: (50~70) % RH
3. Operation Conditions of the Machine  
Set the switch to position “The highest gear”.
4. Test Procedures
  - 1) Get a bacteria slant culture (4~5 generation) which is incubated at 37°C for 24 h, wash the culture from this slant with 10 mL NB, filter the liquid culture by aseptic cotton buds, and dilute this inoculums with NB as appropriate.
  - 2) The equipments are placed in the test chambers respectively, close the door, and open the HEPA filter. Simultaneously operate the environmental control devices until the experimental cabin temperature to be (20~25)°C , relative humidity to be (50~70 )%RH, Turn off the chamber environmental control system.
  - 3) Release microbial aerosol: turn on the microbial aerosol generator, then turn on the ceiling fan, turn off the fan after 5 min, and let stand for 5 min.
  - 4) Original bacteria aerosols collected by six-stage sieve sampler.
  - 5) The test group started the air purifier and sampled after 60 min of action, and the control group also sampled in the corresponding time period.
  - 6) Choose 2 NA plates (the same batch) as the negative control, and culture them on the same condition with the samples.
  - 7) Run the test three times.

5. Computational Formula

$$\text{Natural decay rate } N_t(\%) = \frac{V_0 - V_t}{V_0} \times 100$$

Where:  $V_0$  = Original Bacteria Count of Control group;  $V_t$  = Bacteria Count after Treatment of Control group .

$$\text{Killing Rate } K_t(\%) = \frac{V_1 \times (1 - N_t) - V_2}{V_1 \times (1 - N_t)} \times 100$$

Where:  $V_1$  = Original Bacteria Count of test group;  $V_2$  = Bacteria Count after Treatment of test group.

\*\*\*To be continued\*\*\*

GZ INSTITUTE OF MICROBIOLOGY  
TEST REPORT

Date Received: Jul. 02, 2020  
Date Analyzed: Jul. 02, 2020

**Test results**

Number of Sample	Test Time (min)	Test Bacteria	Test Number	Control Group			Test Group		Killing Rate $K_t$ (%)
				Original Bacteria Count	Bacteria Count after Treatment	Natural Decay Rate	Original Bacteria Count	Bacteria Count after Treatment	
				$V_0$ (cfu/m <sup>3</sup> )	$V_t$ (cfu/m <sup>3</sup> )	$N_t$ (%)	$V_1$ (cfu/m <sup>3</sup> )	$V_2$ (cfu/m <sup>3</sup> )	
KJ20202504-1	60	<i>Staphylococcus aureus</i>	1	$1.22 \times 10^5$	$9.68 \times 10^4$	20.66	$1.36 \times 10^5$	7	99.99
			2	$1.27 \times 10^5$	$1.03 \times 10^5$	18.90	$1.34 \times 10^5$	7	99.99
			3	$1.39 \times 10^5$	$1.11 \times 10^5$	20.14	$1.45 \times 10^5$	7	99.99
		<i>Escherichia coli</i>	1	$1.19 \times 10^5$	$7.99 \times 10^4$	32.86	$1.25 \times 10^5$	7	99.99
			2	$1.14 \times 10^5$	$7.52 \times 10^4$	34.04	$1.10 \times 10^5$	7	99.99
			3	$1.30 \times 10^5$	$8.86 \times 10^4$	31.85	$1.41 \times 10^5$	7	99.99
		<i>Klebsiella pneumoniae</i>	1	$1.24 \times 10^5$	$9.06 \times 10^4$	26.94	$1.20 \times 10^5$	7	99.99
			2	$1.17 \times 10^5$	$8.35 \times 10^4$	28.63	$1.33 \times 10^5$	7	99.99
			3	$1.08 \times 10^5$	$7.87 \times 10^4$	27.13	$1.29 \times 10^5$	7	99.99

Note: The negative control group was sterile growth.

\*\*\*To be continued\*\*\*

GZ INSTITUTE OF MICROBIOLOGY

# Clean Air Delivery

## TEST REPORT

Report Number	KY20200550
Name of Sample	Air Purifier
Applicant	Bryant Medical LTD

GZ INSTITUTE OF MICROBIOLOGY  
**TEST REPORT**

Date Received: Jul. 02, 2020  
Date Analyzed: Jul. 31, 2020

Name of Sample	Air Purifier	Source of Sample	Delivery
Applicant	Bryant Medical LTD	Client	Chen Meifang
Manufacturer	---	Brand	dentair®
Type and Specification	FOZKYGB-03	Quantity of Sample	1PC
Date of Production	---	State of Sample	Machine
Batch Number	---	Packing of Sample	In box
Sample Picture			
Standard and Methods	Referring to ANSI/AHAM AC-1-2019 Method for Measuring Performance of Portable Household Electric Room Air Cleaners		
Items of Analysis	CADR (Pollen)		
Remarks	---		

\*\*\*To be continued\*\*\*

GZ INSTITUTE OF MICROBIOLOGY  
**TEST REPORT**

Date Received: Jul. 02, 2020  
Date Analyzed: Jul. 31, 2020

**Method for Measuring Clean Air Delivery Rate of Pollen:**

1. Test Object  
Particulate (5~11μm)
2. Test Conditions:
  - 1) Environment temperature: (21 ± 3)°C
  - 2) Environment humidity: (40±5) %RH
3. Test Equipment  
Test chamber (30m<sup>3</sup>), Aerosol Spectrometer (TSI 3340), Aerosol Diluter (TSI 3302A), Fluidized Bed Aerosol Generator (TSI 3400A)
4. Operational Conditions of the Machine  
Set the switch to position "The highest gear".
5. Test Procedure
  - 1) Place the air cleaner to be tested in the test chamber in accordance with standard request and set the air cleaner controls to the conditions for test. Test for proper operation, then turn off the air cleaner.
  - 2) Using the test chamber HEPA filter, allow the test chamber air to clean until the background concentration in the size range of (5~11μm) to reaches a concentration of less than 0.03 particles/cm<sup>3</sup>, Simultaneously operate the environmental control devices until the test chamber conditions.
  - 3) Dust is generated in the test chamber by connecting the dust generator, and the pollen generation stops when the pollen concentration reaches (4~9 particles/cm<sup>3</sup>).
  - 4) After the pollen reaches the concentration, the fan will continue mixing for 1min, then turn off and rest for 1min, and confirm again whether the pollen concentration is up to the standard.
  - 5) The air cleaner was turned on for t=0min. Starting from 0min, pollen concentration data were collected every 1min by Aerosol Spectrometer for 10min.
  - 6) Turn off the air cleaner and record the temperature and humidity during the test.
  - 7) Test the natural decay according to the steps 1)~6), except that the air cleaner is unoperated.

6. Computational Formula

$$CADR = (k_e - k_n) \times V$$


Where:  $k_e$  = total decay constant;  $k_n$  = natural decay constant;  $V$  = volume of the test chamber, m<sup>3</sup>

**Test Results**

Number of Sample	Pollutant	Natural Decay Constant $k_n$ (min <sup>-1</sup> )	CADR (ft <sup>3</sup> /min)
KY20200550-1	Pollen	0.14298	353.9

Note: 1ft<sup>3</sup>=0.0283m<sup>3</sup>, 353.9ft<sup>3</sup>/min=600.92m<sup>3</sup>/h.

\*\*\*End of report\*\*\*

Editor  Checker  Issuer  Date Reported 2020.8.10

## Statements

1. The report would be invalid under the following conditions: altered, added, deleted, copied, without the special seal for inspection or signatures by approver.
2. For the received sample, the sample information in the report is claimed by the applicant, the inspection unit is not responsible for its authenticity. The report is responsibility for the received sample only.
3. If there is any objection to the inspection report, it should be presented to the inspection unit within 15 working days from the issuance date, otherwise the report shall be deemed as having been accepted. Microbiological item is not subjected to retest.
4. The report and the name of the inspection unit shall not be used for product labels, advertisements, awards and merchandise publicity.。
5. The items marked with “\*” in the report are not accredited by CNAS or CMA, The items marked with “#” are accredited by CNAS, The items marked with “+” are accredited by CMA.
6. The test data and results of items which are not accredited by CMA, only used as scientific research, teaching or internal quality control.
7. Any ambiguity by the language which used in the report, the Chinese shall prevail.

*Contact Address, NO.1Jiantashan Road, Huangpu District, Guangzhou City, Guangdong Province*

*Test Address, (only fill in when it's different from the contact address)*

*Postal Code, 510663*

*Tel., (8620)61302671*

*URL, <http://www.ggtest.com.cn>*

GZ INSTITUTE OF MICROBIOLOGY

# Sterilization Rate

## TEST REPORT

Report Number	KY20200548
Name of Sample	Air Purifier
Applicant	Bryant Medical LTD

GZ INSTITUTE OF MICROBIOLOGY  
**TEST REPORT**

Date Received: Jul. 02, 2020

Date Analyzed: Jul. 09, 2020

Name of Sample	Air Purifier	Source of Sample	Delivery
Applicant	Bryant Medical LTD	Client	Chen Meifang
Manufacturer	---	Brand	dentair®
Type and Specification	FOZKYGB-03	Quantity of Sample	1 Set (3 PCS)
Date of Production	---	State of Sample	Machine
Batch Number	---	Packing of Sample	In box
Sample Picture			
Standard and Methods	<ol style="list-style-type: none"> <li>1. Referring to GB/T 18801-2015 Air cleaner</li> <li>2. Referring to &lt;Technical Standard For Disinfection&gt; 2002-2.1.3 Air disinfection effect evaluation test</li> </ol>		
Items of Analysis	Removal Rate ( <i>Influenza A virus A/PR8/34 H1N1</i> )		
Remarks	---		

\*\*\*To be continued\*\*\*

GZ INSTITUTE OF MICROBIOLOGY  
TEST REPORT

Date Received: Jul. 02, 2020

Date Analyzed: Jul. 09, 2020

**Test Method for Purification Effect of Airborne Virus Aerosols**

1. Test Equipment
  - 1) Strain: *Influenza A virus A/PR8/34 H1N1*
  - 2) Cells: MDCK
2. Test Conditions
  - 1) Environment temperature: (23~25) °C
  - 2) Environment relative humidity: (50~60) %
  - 3) Test time: 60 min
  - 4) The volume of the test chamber: 30 m<sup>3</sup>
  - 5) Machine setting: "The highest gear".

**Test Results**

Virus	Test Time (min)	Test Number	Virus Titer of Control Group			Virus Titer of Test Group		Removal Rate (%)
			Original Concentration (TCID <sub>50</sub> /m <sup>3</sup> )	Final Concentration (TCID <sub>50</sub> /m <sup>3</sup> )	Natural Decay Rate (%)	Original Concentration (TCID <sub>50</sub> /m <sup>3</sup> )	Final Concentration (TCID <sub>50</sub> /m <sup>3</sup> )	
A/PR8/34 (H1N1)	60	1	3.69×10 <sup>6</sup>	7.03×10 <sup>5</sup>	80.9	5.46×10 <sup>6</sup>	/	≥99.9
		2	2.49×10 <sup>6</sup>	5.85×10 <sup>5</sup>	76.5	1.17×10 <sup>6</sup>	/	≥99.9
		3	7.89×10 <sup>5</sup>	1.98×10 <sup>5</sup>	74.9	3.69×10 <sup>6</sup>	/	≥99.9

**Note:** "/" means not detected.**\*\*\* End of report\*\*\***

Editor

Checker

Issuer

Date Reported

## Statements

1. The report would be invalid under the following conditions: altered, added, deleted, copied, without the special seal for inspection or signatures by approver.
2. For the received sample, the sample information in the report is claimed by the applicant, the inspection unit is not responsible for its authenticity. The report is responsibility for the received sample only.
3. If there is any objection to the inspection report, it should be presented to the inspection unit within 15 working days from the issuance date, otherwise the report shall be deemed as having been accepted. Microbiological item is not subjected to retest.
4. The report and the name of the inspection unit shall not be used for product labels, advertisements, awards and merchandise publicity.。
5. The items marked with “\*” in the report are not accredited by CNAS or CMA, The items marked with “#” are accredited by CNAS, The items marked with “+” are accredited by CMA.
6. The test data and results of items which are not accredited by CMA, only used as scientific research, teaching or internal quality control.
7. Any ambiguity by the language which used in the report, the Chinese shall prevail.

*Contact Address, NO.1Jiantashan Road, Huangpu District, Guangzhou City, Guangdong Province*

*Test Address, (only fill in when it's different from the contact address)*

*Postal Code, 510663*

*Tel., (8620)61302671*

*URL, <http://www.ggtest.com.cn>*



---

  +44 (0)1932 320064 |  +44 (0)7990 113723

[dentair.co.uk](https://www.dentair.co.uk)

Bryant Dental, The Barns, Hilltop Farm, Lyne Lane, Lyne, KT16 0AW  
DentAir® and Bryant Dental® are registered trademarks in the United Kingdom and Australia.  
Registered in England & Wales No. 10396020. Registered in Australia No. 634 752 980

EN2\_0/TRDAJ20 Aug 2020